

# Spectroscopy: Optical and Electronic

Spectroscopy is the use of light, sound or particle emission to study matter. It originated through the study of visible light dispersed, according to its wavelength, by a prism. Later, the concept was expanded greatly to comprise any interaction with radiative energy as a function of its wavelength or frequency.

The data that is obtained from spectroscopy is called a spectrum. A spectrum can be used to obtain information about the specimen's energy states, its atomic or molecular arrangements, and related processes.


# Nature of the interaction between radiation and material

Types of spectroscopy can be distinguished by the nature of the interaction between the probing radiation and the material.


- ◆ Elastic Scattering: incident radiation reflected or transmitted by a material without energy loss.
- ◆ Inelastic Scattering: involving an exchange of energy between the radiation and the matter that shifts the wavelength of the scattered radiation.
- ◆ Refraction: the ability of a medium to impede or slow the transmittance of energy.
- ◆ Absorption: measuring the fraction of incident energy transmitted through the material.
- ◆ Emission: radiative energy released by the material.
- ◆ Resonance: radiative energy couples two quantum states of the material in a coherent interaction.

Spectroscopy originates from the dispersion of light through a prism.

Prism




Grating



$$\begin{aligned}\Delta s &= d (\sin\alpha - \sin\beta) \\ &= m\lambda\end{aligned}$$

# Energy levels and bands







# Electronic and vibrational levels

## Born-Oppenheimer Approximation

$$\Psi_{\text{molecule}}(\vec{r}_i, \vec{R}_j) = \Psi_{\text{electrons}}(\vec{r}_i, \vec{R}_j) \Psi_{\text{nuclei}}(\vec{R}_j)$$



# Transition between two quantum states

## Fermi's golden rule

Transition rate:

$$T_{i \rightarrow f} = \frac{2\pi}{\hbar} |\langle f | H' | i \rangle|^2 \rho,$$

Transition probability


Density of final states

Transition probability is determined by the wavefunction overlapping of initial and final states, and dictated by the selection rules reflecting the nature of interaction  $H'$ .




# Various spectroscopic methods

Spectroscopy is the use of light, sound or particle emission to study matter.




# Laser

The term "laser" originated as an acronym for "light amplification by stimulated emission of radiation".



1. Gain medium
2. Laser pumping energy
3. High reflector
4. Output coupler
5. Laser beam



$$E_2 - E_1 = \Delta E = h\nu$$

# Types of Lasers

Lasers are commonly designated by the type of lasing material employed, which can be a solid, gas, liquid or semiconductor.

**Solid-state lasers** have lasing material distributed in a solid matrix (such as the ruby or neodymium:yttrium-aluminum garnet "Yag" lasers). The neodymium-Yag laser emits infrared light at 1,064 nm.


**Gas lasers** (helium and helium-neon, HeNe, are the most common gas lasers) have a primary output of visible red light. CO<sub>2</sub> lasers emit energy in the far-infrared, and are used for cutting hard materials.

**Excimer lasers** use reactive gases, such as chlorine and fluorine, mixed with inert gases such as argon, krypton or xenon. When electrically stimulated, a pseudo molecule (dimer) is produced. When lased, the dimer produces light in the ultraviolet range.

**Dye lasers** use complex organic dyes, such as rhodamine 6G, in liquid solution or suspension as lasing media. They are tunable over a broad range of wavelengths.


**Semiconductor lasers**, sometimes called diode lasers, are not solid-state lasers. These electronic devices are generally very small and use low power. They may be built into larger arrays, such as the writing source in some laser printers or CD players.

# Interaction of Electron and E Field




- Electrons response to the E field.
- Electrons  $\Rightarrow$  Polarizability  $\Rightarrow$  Susceptibility  $\Rightarrow$  Dielectric function  $\Rightarrow$  Refraction index.
- Resonance will result in absorption and dispersion

# Re-emission and Resonance




- Nonresonance  $\xrightarrow{1-10 \text{ fs}}$  re-emission (**scattering**)
- Resonance : absorption  $\xrightarrow[0.1-100 \text{ ns}]{} \text{re-emission}$  (**photoluminescence**)

# Resonance







# Energy Diagram




# Fluorescence Spectroscopy

Fluorescence spectroscopy, usually using ultraviolet light analyzes fluorescence from a sample. The beam of light excites the electrons in molecules of certain compounds and causes them to emit light; typically, but not necessarily, visible light.




**Jablonski Energy Diagram**




# Bohr model of hydrogen atom

$$E = -\frac{Z^2 m e^4}{8 n^2 h^2 \epsilon_0^2} = \frac{-13.6 Z^2}{n^2} \text{ eV} \qquad r = \frac{n^2 h^2 \epsilon_0}{Z \pi m e^2} = \frac{n^2 a_0}{Z}$$

$$a_0 = 0.0529 \text{ nm} = \text{Bohr radius}$$



# Electron Transitions



A downward transition involves emission of a photon of energy:

$$E_{\text{photon}} = h\nu = E_2 - E_1$$




Given the expression for the energies of the hydrogen electron states:

$$h\nu = \frac{2\pi^2 me^4}{h^2} \left[ \frac{1}{n_1^2} - \frac{1}{n_2^2} \right] = -13.6 \left[ \frac{1}{n_1^2} - \frac{1}{n_2^2} \right] \text{eV}$$

**Fermi's golden rule** is a way to calculate the transition rate (probability of transition per unit time) between two eigenstates

$$T_{i \rightarrow f} = \frac{2\pi}{\hbar} |\langle f | H' | i \rangle|^2 \rho,$$

# Absorption spectra of atomic hydrogen




# Infrared Spectroscopy

Infrared spectroscopy exploits the fact that molecules absorb specific frequencies that are characteristic of their structure. These absorptions are resonant frequencies.

The infrared portion of the electromagnetic spectrum is usually divided into three regions; the near-, mid- and far- infrared, named for their relation to the visible spectrum. The higher-energy near-IR, approximately  $14000\text{--}4000\text{ cm}^{-1}$  ( $0.8\text{--}2.5\text{ }\mu\text{m}$  wavelength) can excite harmonic vibrations. The mid-infrared, approximately  $4000\text{--}400\text{ cm}^{-1}$  ( $2.5\text{--}25\text{ }\mu\text{m}$ ) may be used to study the fundamental vibrations and associated rotational-vibrational structure. The far-infrared, approximately  $400\text{--}10\text{ cm}^{-1}$  ( $25\text{--}1000\text{ }\mu\text{m}$ ), lying adjacent to the microwave region, has low energy and may be used for rotational spectroscopy.




# Absorption spectra of molecular hydrogen



Zhi Ping Zhong et al., Phys. Rev. A **60**, 236 (1999)


# Raman Spectroscopy


Raman spectroscopy is a spectroscopic technique used to observe vibrational, rotational, and other low-frequency modes in a system. Raman spectroscopy is commonly used in chemistry to provide a fingerprint by which molecules can be identified.








# Raman and Rayleigh Scattering








**Figure 8.19.** Raman spectra of (a) crystalline graphites and (b) noncrystalline, mainly graphitic, carbons. The D band appears near  $1355\text{ cm}^{-1}$  and the G band, near  $1580\text{ cm}^{-1}$ . [From D. S. Knight and W. B. White, *J. Mater. Sci.* 4, 385 (1989).]



# Spectroscopy at nanometer scale

1. Spectroscopy vs. Microscopy
2. Physics and Chemistry of nanomaterials

# Resonance




Microscopy is the science of investigating small objects that are too small for the naked eye. The microscopic study involves revelation of the structure and morphology of the matter under investigation.

Spectroscopy is the study of the interaction between matter and radiated energy. With the mechanism of “resonance”, the characteristic nature of the matter can be probed.

**Both microscopic and spectroscopic techniques are essential for nanoscience research.**

# One dimensional size effect




$$\Psi(x) = \begin{cases} \sin(n\pi x/a), & n \text{ even} \\ \cos(n\pi x/a), & n \text{ odd} \end{cases}$$


$$E = n^2 \pi^2 \hbar^2 / 2ma^2, \quad n = 1, 2, 3, \dots$$



# Size effect



## **Au nanoparticle as an example**



← 10 nm →


$$E_F = (\hbar^2/2m) (3\pi^2n)^{2/3}$$

$$g(E_F) = (3/2) (n/E_F)$$

$$\delta = 2/[g(E_F)V] = (4/3) (E_F/N)$$




Number of valence electrons (N) contained in the particles is roughly 40,000. Assume the Fermi energy ( $E_F$ ) is about 7 eV for Au, then

$$\delta \sim 0.22 \text{ meV} \sim 2.5 \text{ K}$$




**Ultraviolet photoemission** spectra of ionized copper clusters  $\text{Cu}_N^-$  ranging in size from  $N$  of 1 to 410 show the energy distribution versus binding energy of photoemitted electrons. These photoemission patterns show the evolution of the 3d band of Cu as a function of cluster size. As the cluster size increases, the electron affinity approaches the value of the bulk metal work function. (Adapted from ref. 10.) **Figure 5**

# Semiconductor quantum dots




(Reproduced from Quantum Dot Co.)






# Optical properties of nanoparticles (in the infrared range)




(1) Broad-band absorption:  
Due mainly to the increased  
normal modes at the surface.

(2) Blue shift:  
Due mainly to the bond shortening  
resulted from surface tension or  
phonon confinement.



**Figure 8.11.** Raman spectra of Ge nanocrystallites produced by gas condensation showing the broadening and shift to lower wavenumbers as the particle size decreases. The lowest spectrum was obtained from ground bulk germanium. [From S. Hayashi, M. Ito, and H. Kanamori, *Solid State Commun.* **44**, 75 (1982).]

# Optical properties of nanoparticles (in the visible light range)



## (1) Blue shift:

Due mainly to the energy-gap widening because of the size effect.

## (2) Red shift:

Bond shortening resulted from surface tension causes more overlap between neighboring electron wavefunctions. Valence bands will be broadened and the gap becomes narrower.

## (3) Enhanced exciton absorption:

Due mainly to the increased probability of exciton formation because of the confining effect.


# Scanning Tunneling Spectroscopy

## 1. Barrier Height Imaging

Up to now homogeneous surfaces were considered. If there is an inhomogeneous compound in the surface the work function will be inhomogeneous as well. This alters the local barrier height. Differentiation of tunneling current yields


$$\frac{d(\ln I)}{ds} \propto \sqrt{\Phi}$$

Thus the work function can directly be measured by varying the tip-sample distance, which can be done by modulating the current with the feedback turned on.






# STM Images of Si(111)-(7×7)



Empty-state image





Filled-state image


# Electronic Structures at Surfaces

## Empty-State Imaging


### Not Tunneling



### Tunneling



## Filled-State Imaging




## 2. dI/dV imaging

If the matrix element and the density of states of the tip is nearly constant, the tunneling current can be estimated to

$$I \propto \int_0^{eV} \rho_{sa}(E_F - eV + \varepsilon) d\varepsilon$$

Differentiation yields the density of state:


$$\frac{dI}{dV} \propto \rho_{sa}(E_F - eV)$$



The mapping of surface density of states can be deduced by


- Modulation of the bias voltage (dI/dV imaging):  
The tip is scanned in the constant current mode to give a constant distance to the sample. A dither voltage of  $\sim 1\text{ kHz}$  is added to the bias voltage while the feedback loop remains active. A lock-in technique is employed to obtain the current change at the dither frequency.
- Current-Imaging Tunneling Spectroscopy (CITS):  
The tip is scanned in the constant current mode to give a constant distance to the sample. At each point the feedback loop is disabled and a current-voltage curve (I-V curve) is recorded.

# STS of Si(111)-(7x7)



(a)

Science 234, 304 (1986).




(b)


# STS of Si(111)-(7x7)

topograph


+2V





-0.8V



-0.35V





-1.8V



1. Science **234**, 304-309 (1986).
2. Phys. Rev. Lett. **56**, 1972-1975 (1986).

# Electronic Structure of Single-wall Nanotubes







Nature **391**, 59 (1998).

# Quantum size effect


$\lambda$  = de Broglie wavelength of electron

$a$  = thickness of metal film







## Pb islands on the IC Pb/Si(111)



# Spectra for Pb Films







C.M. Wei and M.Y. Chou

$$d_0 = 2.85 \text{ \AA} \quad \lambda_F = 3.94 \text{ \AA}$$



$$2d_0 \approx 3(\lambda_F/2)$$

# Inelastic Tunneling

## Elastic vs. Inelastic Tunneling




# Single Molecule Vibrational Spectroscopy and Microscopy



B.C. Stipe, M.A. Rezaei, and W. Ho,  
Science **280**, 1732-1735 (1998).





# Quantum corral



D.M. Eigler, IBM, Amaden

# Artificial atom




# Quasiparticle Interference (QPI)







In an ideal metal, the Landau quasiparticle eigenstates are Bloch wavefunctions characterized by wavevector  $\mathbf{k}$  and energy  $\varepsilon$ . Their dispersion relation,  $\varepsilon(\mathbf{k})$ , can be measured by real space imaging techniques, such as scanning tunneling microscopy (STM). This is because the local density of states LDOS ( $E$ ) spectrum at a single location  $\mathbf{r}$  is related to the  $\mathbf{k}$ -space eigenstates  $\Psi_{\mathbf{k}}(\mathbf{r})$  by

$$\text{LDOS}(E, \mathbf{r}) \propto |\Psi_{\mathbf{k}}(\mathbf{r})|^2 \delta(E - \varepsilon(\mathbf{k}))$$

Sources of disorder such as impurities or crystal defects cause elastic scattering which mixes eigenstates of different  $\mathbf{k}$  but the same  $\varepsilon(\mathbf{k})$ . In other words, elastic scattering mixes states that are located on the same quasiparticle contour of constant energy (CCE) in  $\mathbf{k}$ -space. When scattering mixes states  $\mathbf{k}_1$  and  $\mathbf{k}_2$ , the result is a standing wave in the quasiparticle wavefunction  $\Psi_{\mathbf{k}}$  of wavevector  $\mathbf{q} = (\mathbf{k}_1 - \mathbf{k}_2)/2$ . Since LDOS is proportional to the norm of the quasiparticle wavefunction  $|\Psi_{\mathbf{k}}|^2$ , the LDOS will contain an interference pattern with wavevector  $\mathbf{q} = \mathbf{k}_1 - \mathbf{k}_2$ , or wavelength  $\lambda = 2\pi/q$ .







(a) 12 meV  $|FT(DOS)|$


(b) 16 meV  $|FT(DOS)|$

(c) 22 meV  $|FT(DOS)|$




# Optical methods with sub-wavelength spatial resolution

## Aperture near-field probes



Spatial resolution	100 nm
Time resolution	100 fs - cw
Spectral resolution (cw)	30 $\mu$ eV
Sample temperatures	10 - 300 K


## Apertureless scattering




10 nm
10 fs - cw
several meV
300 K

# Recent development: AFM-IR system

AFM-IR can perform IR spectroscopic chemical identification with sub-100 nm spatial resolution



Scheme of the AFM-IR setup. The AFM cantilever ring-down amplitude plotted as a function of laser excitation wavelength produces the IR spectrum.



(a) AFM topography picture of the bacterium; the position of the tip is indicated in blue. (b) FT-IR spectrum; the bacterium absorption spectrum is drawn in green, and the wavenumber of the CLIO laser is indicated by the red arrow. (c) Oscillations recorded by the four-quadrant detector (in red) as function of time superposed on the CLIO pulse laser (blue).

Alexandre Dazzi et al., APPLIED SPECTROSCOPY OA 66, 1365 (2012)


# Surface-enhanced Raman spectroscopy (SERS)

**SERS** is a surface-sensitive technique that enhances Raman scattering by molecules adsorbed on rough metal surfaces or by nanostructures such as plasmonic-magnetic silica nanotubes. The enhancement factor (EF) can be as much as  $10^{10}$  to  $10^{11}$ , which means the technique may detect single molecules.

The electromagnetic theory proposes the excitation of localized surface plasmons ( $EF \propto E^4$ ), while the chemical theory proposes the formation of charge-transfer complexes. The chemical theory applies only for species that have formed a chemical bond with the surface, and the electromagnetic theory of enhancement can be applied regardless of the molecule being studied.


# SERS

## Surface Enhanced Raman Scattering




- SERS is a surface sensitive technique.
- Huge enhancement on rough metal surfaces.
- The enhancement factor can be as much as  $10^{14} - 10^{15}$ .
- High sensitivity possibly for single molecule detection.

# Tip Enhancement






- LSP results in enhancement of EM field.
- Similar to surface enhanced Raman scattering (SERS)
- Provides location dependent tip enhanced Raman spectrum.

# Tip-enhanced Raman spectroscopy




Topography

1594  $\text{cm}^{-1}$  vibration









# STM TERS Imaging of CNT



40 x 30 px




Simultaneous STM and Raman imaging.




C. Chen, N. Hayazawa, and S. Kawata, Nature Comm. 5, 3312 (2014)



# Light scattering from a metallic nanotip



$r=10-20$  nm




Scatterer dimension  $a=2r \ll \lambda$   
(Mie scattering)

Electric field enhancement for field vector parallel to tip axis:

- lightning rod effect (field singularity at tip apex)
- surface plasmon excitation



Field enhancement factor 10 – 50




*L. Novotny et al., Phys. Rev. Lett. 79, 645 (1997),  
Annu. Rev. Phys. Chem. 57, 303 (2006)*

# Far Field Super Resolution

- Airy disk (point spread function)





- From a single emitter.  
(Fluorescence protein or dye)
- Localize the center of PSF.





# Localization - PALM

## Photo-Activated Localization Microscopy





# Localization - STORM

## Stochastic Optical Reconstruction Microscopy





# X-ray Photoelectron Spectroscopy (XPS)



$$E_i = E_{h\nu} - (E_k + \phi)$$

# X-ray source




# Concentric Hemispherical Analyzer (CHA)




$$\Delta E/E_0 = s/R_0$$

s: mean slit width;  $R_0$ : mean radius





# Angle-resolved photoemission spectroscopy (ARPES)



We need:

binding energy -  $E_b$

initial momentum -  $k^i$

$$E_b = E - h\nu + W$$

$$k_{||}^i = k_{||}^f = \sqrt{2mE/h^2} \sin\theta$$

$$k_{\perp}^i = k_{\perp}^f - G = \sqrt{2mE/h^2} \cos\theta - G$$

